



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore - 560 068.

Ph: 080-61754680 Fax: 080 - 61754693E-mail:deandirectortodc@gmail.com

Website: www.theoxford.edu

Logbook or other records regarding maintenances work

Index

Sl. No.	Particulars	Page No.
1.	Logbook regarding maintenances works	2 - 161
2.	Installation report for RO plant	162 - 165
3.	Installation report for ST plant	166 - 184
4.	Maintenance expenses	185 - 189

Dean and Director

The Oxford Dental College, Bommanahalli
Hosur Road - Bangalore - 560 068

22

(146)

2021

Date Name of the Dept Nature of the complaint

22/3/2021

Ortho PG

Dental light
not working in
one Number of dental
Chair

12/4/2021

conservative leg.

Auto clare case
made not working

15/7/2021

(conservative
U.G.)

Chair No 14
Back rest
movement not
working in
MRB dental chair

147)

2021

work report -

Staffle son

Attended the complaint,
in dental chair,
Halogen bulb is, tested
Replaced the new bulb
against old.
Now checked and found
in working condition.

~~22/3/2021~~ 22/3/2021 J

Attended the complaint,
in Auto clave PCB problem. (i.e after
service from Continer) re installed
the PCB and checked with
one cycle, and found in working
condition (BOM Invoice No 048)

~~12/4/2021~~ 12/4/2021
Date 8/4/2021
checked
12/4/2021

Attended the complaint,
in dental chair; problem
in the motor. Replaced the
new back rest motor
against old. Now checked
and found in working condition

~~15/7/2021~~
15/7/2021

15/7/2021

(148)

2021

Date	Name of the person	Nature of the complaint
<u>19/7/21</u>	<u>Pedo. P.G.</u>	C-79 Auto clare not working

<u>24/7/21</u>	<u>Ponbo P.G.</u>	water leakage in three way syringe and both buttons of the three way not working <u>Mr. Prigenba P.G.</u>
----------------	-------------------	--------------------------------------------------------------------------------------------------------------

<u>23/7/21</u>	<u>Pedo P.G.</u>	Water leakage at ^{Head} Phantom headrolley.
----------------	------------------	------------------------------------------------------

2021

(149)

work. or part

staff sign

Attended the complaint
in auto clove, Thermostat
of wet coil are burnt
Replaced the new Heating coil
of Thermostat against defective
checked with one cycle
as it is in working condition

19/7/2021

(Invoice No 500)
dated 13/7/21

19/7/2021

Attended the complaint
in dental chair three way/syringe
problem replaced the new
Three way syringe against
defective. Now checked
found it is in the working
condition

19/7/2021

As

Attended the complaint
in the ^{two} trolleys. was
problem in the W.R.V.
Replaced the new W.R.V.
against for the two number
of Phenoxan trolleys.

Now checked found
that both trolleys worked
properly.

28/7/2021

(15)

2021

Date

Name of student

Nature of the defect

27/7/21

Ortho P.G.

Water not coming in the three way ^{sample}
(Dr. Sneha. P.G.)

30/7/21

Pedo P.G.

Water leakage at airator pond.
(in char No. 15)

11/8/2021

(Perio P.G.)

C-79 Autoclave (B)
not working

251

2021

Work report

Staff's sign

Attended the complaint in the dental chair, problem in the three way syringe

Replaced the new three way syringe against defective one. Now checked and found it is in good working condition.

21/7/21

~~Signature~~
22/7/21

Attended the complaint in the dental chair, problem in the W.R.V.

Replaced the new W.R.V. against old. Now checked and found it is in the working condition.

23/7/21

~~Signature~~
30/7/21

Attended the complaint in the autoclave problem in the wet coil. Replaced the new coil against burnt one, and checked with one cycle found that it is in the working condition.

(invoice NO 771/9/8/21)

~~Signature~~
11/8/2021

(152)

2021

Date	Name of the dept	Nature of the complaint
7/10/21	Pedo.	Two Member of dental chair were Problem in the section motor, getting more sound.

<u>5/10/21</u> (conservative pg)		x-ray head cone need to be replace
----------------------------------	--	------------------------------------

<u>16/11/21</u>	<u>Pedo P.G.</u>	2 Nos of mbr dental chair up down movement not working
-----------------	------------------	--------------------------------------------------------

<u>16/11/21</u>	<u>Pedo P.G.</u>	one Member of Chaurandi dental chair up down movement not working.
-----------------	------------------	--------------------------------------------------------------------

2021

(153)

work sent

staff's sign

Attended the complaint in the Suction motor problem in the bearing, after replacement of bearings, Two motors were ~~file~~ checked found it is in the wrong condition

~~SP~~ (Bill No. 1530)
 done 7/10/21
 M. Deff Review by work. 8/10/2021

Attended the complaint in X-ray head, cone was damaged. Replaced the new cone against old.

(Bill No. 207 4/10/21)
 (S.B. X-ray systems)

Attended the complaint in the dental chair problem in the key board, replaced the new key boards against old. Now checked and found that both chair were working properly

(Bill No. 1629/11/11/21)
 ~~SP~~
 16/11/21

Attended the complaint in the dental chair problem in the PCB. After repairs from confident. Now installed & checked found in the working condition

(Bill No. 1625/10/11/21)
 ~~SP~~
 16/11/21

154

Date
24/11/2021

Name of the dept
Peco PG,

Name of the
component

char No (15)

m.m.g. way
holder need to be
replace

27/11/21

Ortho (P.G, lab)

Double disc
model commes
not working

3/12/21

(Peco U.G)

one Number dental
char problem in the
air leakage
char No (U.G. 5)

4/12/21

(Peco P.G)

char No (P.G. 1)
Aircraft not working

255

2021

work report

Usha Gold
Page No. 201
Date / / 2021

Staff's
Sign.

Attended the complaint
in chair No (15) problem
in the holder i.e broken
Replaced the new ring chuck
holder against old,

[Signature]
24/11/21

[Signature]
24/11/21

Attended the complaint in the
model trimmer winding is burnt
After repaired, & installed and
checked found in the working
condition.

Bill No (1555) Dated 26/11/21 (Winding
works)

[Signature]
27/11/21

[Signature]
27/11/21

Attended the complaint
in the dental chair, air tube
is damaged. Replaced the
new air tube (3 meter) against
damaged. Now checked and found
in the working condition.

[Signature]
3/12/21

[Signature]
2/12/2021

Attended the complaint in the
dental chair, Problem in the air tube
i.e damaged. Replaced the new air tube
from trolley to baseline box (4 meter). Now
checked and found in the working condition.

[Signature]
4/12/21

[Signature]
4/12/2021

(156)

2021

Date	Name of the dept	Nature of the complaint
<u>10/12/21</u>	Pedo. P.G	2 Nos of dental chair are not working 1 No of dental chair is Problem water leakage Chair Nos - <u>PG-03</u> / <u>PG-08</u>

(PG.04)

Pedo. UG

one Number of dental chair Problem on the air leakage (UG.15)
 one Number of dental chair Problem in the suction

10/12/21

(conservative UG)

Chair No (13) Problems in the W.R.V
 Chair No (19) air leakage needs to be replaced
 Chair No (20) air tube needs to be replaced
 Chair No (27) low suction tip holder needs to be replaced
 Chair No (23) air tube damaged needs to be replaced

Work report
Attended the complaint
^{for 2nd} in dental chair Problem
in the air tube re damaged
Replaced the new air tube from
baseline bin to trilogy. Total
3mtr of air tube was consumed
Now checked & found in the
working condition.
in one Number of dental chair
Problem in the w.r.v. Replaced
the co.rod against defective
Now no water leakage
10/12/21

10/12/21

Attended the the complaint
in dental chair Problem in the
air tube Replaced the new on the
entry against damaged.
in the another chair have Problem
in the Suction elbow. Replaced the
new Suction elbow against broken on
10/12/21

11/12/21

Attended the complaint in
Chair No (13) Problem in the w.r.v
Replaced new w.r.v against dpt.
Chair No 31, 30, & 23 Problem
in the air tube. Replaced
the air tube against damaged on
each 3mtr ~~total~~ for one chair
total 9 mtr was used.
Chair No (27) Problem in the low Suction
in handles Replaced the same.
11/12/21

158

2021

Order
31/12/2021

Name of the dept

Nature of the
complaint

(ortho P.G.)

left down movement
not working

7/1/2022

(ortho P.G.)

Dental light not
working

(Dr. Muralin PG)

9/3/2022

conservative
(U.G.)

Ruckin is not
working in the
dental chair (change)

17/3/22

(ortho P.G.)

C-79 Auto clave not
working

(159)

2022

Work report

Usha Gold
Page No. 201
Staff's Sign

Attended the Complaint
in the dental chair ^{updown}
hand control switch Problem. 31/12/22
Replaced the new switch
against old Now checked
found into working condition.

~~pep~~
31/12/2021

Attended the Complaint in the
Halogen bulb was fused,
Replaced the new bulb against
old. Now it is into working
condition. 31/12/22

~~pep~~
31/12/2021

Attended the complaint in the
suction, holder Problem
Replaced the new suction holder
against old, now checked and
it is into working condition. 9/3/22

~~pep~~
9/3/22

Attended the Complaint in the
autoclave, wet coil, and thermostat
where burnt. Replaced the
both parts against old.
now checked with one cycle
found into working condition.
invoice NO 2671 dated 10/3/2022

~~pep~~
17/3/2022
17/3/22

2022

<u>Date</u>	<u>Name of the dept.</u>	<u>Nature of the complaint</u>
<u>21/3/22</u>	(Pedo P.G. pre clinical lab)	(water leakage at anoter point)
<u>7/4/2022</u>	(conservative P.G)	(water leakage at another point in Pwja P.G)
<u>16/5/2022</u>	(Perio U.G.)	C-79 Autoclave not working


2022

work report

Usha Gold
Page No. _____
Staff's Sign


Attended the complaint-
in the pre clinical ^{PG} lab, (redo)
Problem in the W.R.V.

Replaced the new W.R.V
against defective. Now checked
and found ~~that~~ it is in the working
condition. ~~NO~~ water leakage



21/3/2022


Attended the complaint
in the dental chair Problem
in the W.R.V. Replaced the
new W.R.V against old
Now checked & found in the
working condition.


14/3/2022

Attended the complaint
in the dental chair Problem
in the heating coil. (i.e burnt)
Replaced the new wet coil

against burnt one. Now checked
with one cycle ^{if it into} working condition

(Invoice ^{Bill No.} No 28) / 6/5/2022
From Sugandh-cum-dain


16/5/22

16/5/2022

2022

Date. Name of the dept. Nature of the complaint
27/6/2022 Podology air leakage in dental chair

27/6/2022 Podology dental light not working in one No of ~~dent~~ chair

28/6/2022 Ortho P.G Friller & arator holder need to be replace

- do - - do - Low section^{HP} holder need to be replace

5/7/2022 Ortho P.G section not working

work report

Attended the complain
in dental chair air tube was
damaged. Replaced the new air tube
against damaged one. Now checked and
found in the working condition
[Signature] 30/6/2022

Attended the complain
in dental chair, Halogen bulb
tube replaced new halogen bulb
against old.
[Signature]

attended the complain
in trilled airtor holder problem
replaced the new airtor holder
against damaged one
[Signature]

28/6/2022

fr. shankar
28/6/22

Replaced the new low suction
tip holder against broken one
[Signature]

28/6/2022

fr. shankar
28/6/22

Attended the complain
in the suction, after service from
sugident, PCB had re installed
and checked found in the working
condition. (invoice No 742. sl. No. 5)

Dated 2/7/2022

5/7/2022

2022

<u>Date</u>	<u>Name of the dept</u>	<u>Nature of the complaint</u>
<u>5/7/2022</u>	<u>Pedo P.G</u>	chair up/down movement not working (Dr. Anushree P.G)
<u>5/7/2022</u>	<u>Conservative P.G</u>	one Number of dental chair problem in the Sensor light and micromat not working
<u>6/7/2022</u>	<u>Perio P.G</u>	dental light working (charan B)
<u>13/7/2022</u>	<u>Ortho P.G</u>	dental chair up/down movement not working

2022

Work of Park

Staff's Sign

Attended the complaint in the chair, ^{Pcb} after serviced from urgent corporation, re installed and checked all in the working condition.
(invoice No. 742. Sl.No. (4) Dated 2/7/2022)

[Signature]
5/7/2022

[Signature]
5/7/2022

Attended the complaint in dental chair ~~sensor~~ problem in sensor from Pcb / after serviced from urgent, re installed and checked both functions were in working condition.

(invoice No 742 Sl.No. 2 and 3 Dated 2/7/2022)

[Signature]
5/7/22

[Signature]
5/7/22

Attended the complaint in the dental chair problem in the Triley cpu Pcb of key board. Replaced the new Triley cpu and key board, now checked and found in the working condition.

(invoice No 742 Sl.No. (1) Dated 2/7/2022)
(Pa-Du AIDA)

[Signature] 6/7/22

[Signature]
7/7/2022

Attended the complaint in the dental chair problem in the up/down motor.

Replaced the new up/down motor against defective. Now checked and found in the working condition. complaint ^{given} on 2/7/22

[Signature]
12/7/22

[Signature]

Dean and Director
The Oxford Dental College, Borenagahalli
Hosur Road, Bengaluru - 560 087