



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore – 560 068.

Ph: 080-61754680 Fax: 080 – 61754693E-mail:deandirectortodc@gmail.com

Website: www.theoxford.edu



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore – 560 068.

Ph: 080-61754680 Fax: 080 – 61754693E-mail:deandirectortodc@gmail.com

Website: www.theoxford.edu

Logbook or other records regarding maintenances work

Index

Sl. No.	Particulars	Page No.
1.	Logbook regarding maintenances works	2 - 161
2.	Installation report for RO plant	162 - 165
3.	Installation report for ST plant	166 - 184
4.	Maintenance expenses	185 - 189

PRINCIPAL

**The Oxford Dental College
Bommanahalli, Hosur Road,
Bangalore - 560 068.**



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore - 560 068.

Ph: 080-61754680 Fax: 080 - 61754693E-mail: deandirectortodc@gmail.com

Website: www.theoxford.edu

<u>27/6/2022</u>	<u>Pemb P.G</u>	dent light not working in one No of dent chair
<u>28/6/2022</u>	<u>Ortho P.G</u>	Fritley A injector holder need to be replace
- do -	- do -	low scrubbing ^{off} holder need to be replace
<u>5/7/2022</u>	<u>Ortho P.G</u>	section not working



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Rommanahalli Hosur Road Bangalore - 560 068

2022

work report

Staff's Sign:

Attended the complain

in dental chair air tube used

damaged. Replaced the new air tube
against damaged one. Now checked and

found in the working condition

30/6/2022

Attended for complain

in dental chair, Halogen bulb

failed replaced new halogen bulb
against old.

attended the complain

in tilted armator holder problem

replaced the new armator holder
against damaged one

28/6/2022

28/6/22

Replaced the new low suction
tip holder against broken one

28/6/2022

28/6/22

Attended the complain

in the suction, after service from

supplier, PCB had re installed

and checked found in the working

condition. (invoice No 742. sl. No. 5)

dated 2/7/2022

5/7/2022

2022

Date	Name of the dept	Nature of the Complaint
5/7/2022	Pedo P.G	Chair up/down movement not working (Dr. Anushree P.G)
5/7/2022	Conservative PG	one Number of dental chair problem in the Sengox light and micromat not working
6/7/2022	Pedo PG	dental light working (Chinnik)
13/7/2022	ortho P.G.	dental chair up/down movement not working



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences, Karnataka & Recognised by Dental Council of India, New Delhi)

Rommanahalli Hosur Road Bangalore - 560 068


2022


Work of Park

Staff (S. Sign)

Attended the complaint in the chair ^{pcb} after serviced from Sugdent Corporation, re installed and checked all in the working condition.


(invoice NO. 742 Sl. NO. (4) Dated 2/7/2022)



5/7/2022


5/7/2022

Attended the Complaint in dental chair sensor problem in sensor form PCB after serviced from Sugdent, re installed and checked both functions were in working condition.

(invoice No 742 Sl. No. 2 and 3 Dated 2/7/2022)


5/7/22



5/7/22

Attended the Complaint in the dental chair problem in the Trolley CPU PCB key board. Problem. Replaced the new Trolley CPU and Key board, Now checked and found in the working condition.

(invoice NO 742 Sl. NO (1) Dated 2/7/2022)

(Pa-Du AIDA)


6/7/22


7/7/2022

Attended the Complaint in the dental chair Problem in the up/down motor.

Replaced the new up/down motor against defective. Now checked & found in the working condition. Complaint given on 12/7/22


13/7/2022





2022

Date	Name of the dept	Nature of the complaint
13/9/22	conservative (P.G.) (Esthetic dept)	Sensu ligul not working

22/9/22	conservative P.G.	char No (C) Dr. Madhura AG spitten coupling needs to be reduce
---------	-------------------	---

22/9/22	Pedio P.G.	char nos AG (CS) Dr. Acelab. P.G. water leakage at anoter point
---------	------------	--

Spoken
complaint
resolved

2022

~~work sheet~~

Staff's Sign

After service from
Surgident. we found the
Senior Rent Pcb. checked
and found in the working condition
invoice NO. 1342 Dated 7th Sept 2022
(Bill amount 2700) from Surgident Corporation.

~~13/9/2022~~

~~15/9/22~~

~~15/9/22~~

attended the complaint in
chair N(5) Spitter Coupling
was broken. Replaced the
new Spitter Coupling against
broken one. Now no water
leakage in the chair.

~~22/9/22~~

~~22/9/22~~

Attended the complaint
in the dental chair problem
in the W.R.V. Replaced the new
W.R.V. against. Old,
checked & found in the working condition
no water leakage

~~22/9/22~~

~~22/9/22~~

2022

Sciences,

<u>Date</u>	<u>Name of the dept</u>	<u>Nature of the Complaint</u>
<u>26/9/22</u>	Conservative (4G)	air rotor not working properly.

<u>5/11/22</u>	Ortho PG	cutty not getting off (suction holder needs to be replace)
----------------	----------	--

<u>24/11/22</u>	(Pedo PG)	chair up down not working. (also attached PG)
-----------------	-----------	---



2022

Work report

Staff Sign

Attended the complaint in chair No 2 problem in the amtor foot control. Replaced the new amtor foot control against defective. Now checked and found it is in working condition

~~26/9/22~~
26/9/22

Replaced
Ppt
26/9/22

attended the complaint in Sucky holder holder problem replaced the new Sucky holder against old, now checked and found in working condition

~~5/11/22~~
5/11/22

Attended the complaint in the dental chair PCB key board problem replaced the new key board against defective one. checked and found in the working condition

(INVOICE NO 1960/22/11/22 sl. No 1 of 2)
from Sufient Corporation,

~~24/11/22~~
24/11/22

for
24/11/22



th Sciences,

pm

2/12/22 - concretive (4.G.) water leakage
at corner point

20/1/23 Perfo P.G. churnu (12)
OP chair up/down
movement not work

1.1.23

Pracette

2022

Usha Gaid
Page No.
Date: 2/12/22

Working Report

Attended the complaint in dental chair Problem in the spliton Coupling (ie broken)
Replaced the new spliton coupling against broken one.

In micromotor PCB Problem, ie repaired from Suspect Component re installed the same

checked & found it is into working condition. ~~ADP~~ (PG) 2/12/22

Attended the complaint in dental chair Problem in the W.R.V. Replaced the new W.R.V against defective one. checked & found it is into working condition no water leakage.

~~ADP~~ (V9) 2/12/22

Attended the complaint in the dental chair Problem in the key board Cable Replaced the same
Now checked & found in the working condition

~~ADP~~
20/1/23

Attended the complaint in automate Processor Problem in the circuit board, so send to JDS, after repaired, re installed the same. Now checked & found it is into working condition. invoice NO JDS/5477/22-23

~~ADP~~ Annually rated 9/Jan/23

Math Sciences,

om

Date	Name of the department	Nature of the Complaint
6/5/2023	Charit Pedo P.G.	Char No (15) Dental light not working.
29/5/2023	Pedo P.G.	1) Char No. 5, 6 Eviction not working 2) Char No (13) & me Scales not working. 3) Char No (15) Dental light not stable.
5/6/2023	Conservative P.G.	Char No (11) Water not coming in orator Point.
6/6/2023	Pedo P.G.	Char No (8) Water not coming in Bio. Optic Scales
10/6/2023	Conservative U.G.	Char No (20) and leakage in dental chair



Date / /

Page

work report

Attended the complaint in chair No (15). 12 volts supply wire problem, rectified & solved the same.

26/5/2023

26/5/23

Attended the complaint in chair No, 5, 6, problem in the suction sensor solved the same.

in chair No (13) emg scaler cord problem rectified & solved the same.

in chair No (15) hinge alignment problem solved.

29/5/2023

Attended the complaint in chair No (11) problem in the booster rectified and solved the same.

5/06/2023

Attended the complaint in chair No (5) problem in the scaler solonide valve rectified and solved the same.

5/06/2023

Attended the complaint in the chair No (20) problem in the air tube (re damaged)

Solved the same. Now it is the working conditions.

10/6/23.



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences,
Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore – 560 068.

Ph: 080-61754680 Fax : 080 – 61754693E-mail:deandirectortodc@gmail.com

Website: www.theoxford.edu

Date / /

Page Staff Sign

Work done report

Attended the Complaint
 in class No (2) Problem
 in Control LED of class PCB,
 both parts replaced against
 old. Now checked & found
 it is in the working condition
 (invoice No 639/6th June 2023)
 from Sargident Corporation
 .12/6/2023

[Signature]

Attended the Complaint in
 the class No (18), class Control
 Problem had been PCB. after
 received from Sargident Corporation
 & installed the same. Now
 checked & found in the working condition
 invoice No. 639/6/6/2023.
 from Sargident Corporation.
 .12/6/2023.

Installed & working for 12/6/23.

Attended the Complaint
 in (15) Problem in the Switch
 solved the same.

Attended the Complaint
 in identical class problem in the
 air tubes solved the same

 16/6/23

Attended the Complaint
 in class (11) Problem in the
 section holder solved the
 same
 16/6/23



CHILDREN'S EDUCATION SOCIETY (Regd.)

THE OXFORD DENTAL COLLEGE

(Recognized by the Govt. of Karnataka, Affiliated to Rajiv Gandhi University of Health Sciences, Karnataka & Recognised by Dental Council of India, New Delhi)

Bommanahalli, Hosur Road, Bangalore – 560 068.

Ph: 080-61754680 Fax : 080 – 61754693E-mail:deandirector@odc@gmail.com

Website: www.theoxford.edu

in char No (9) - 2 Nos of Three way
Syringe replaced,

in char No (6) ^{work} 01 No Three way and Light Switch
replaced,

in char No (2) one No. Three way
and airator holder had been replaced

in char No (3) airator holder had
replaced.

in char No (8) airator holder had replaced

in char No (10) - 2 Nos of Three way had been
replaced

in char No (11) 1 No. Three way and airator
holder. had been replaced,

one (1) No. of three way Syringe
had replaced against old.

Attended the complaint in (14)
one Number of Three way Syringe
replaced against old.

in char No. (4) Replaced the
New dust cover, three way Syringe and
3 metr. air tube (big) against old.

Now checked and found in the working
condition.

Attended to complaint in the dental
char air tube had damaged replaced the
new air tube i.e 3 metr (big) against old,

Spares - air tube
Three way Syringe - 10.
Airator holder
Switch cap
dust cover

in chair No (5) - 2 Nos of Three way syringe replaced.

in chair No (6) - 01 No Three way and Light switch replaced.

in chair No (7) one No. Three way and motor holder has been replaced.

in chair No (8) motor holder has been replaced.

in chair No (9) motor holder has been replaced.

in chair No (10) - 2 Nos of Three way has been replaced.

in chair No (11) 1 No. Three way and motor holder has been replaced.


in chair No (12) 01 No. of three way syringe has replaced against old.

attended the complaint in (14) one Number of Three way syringe replaced against old.

in chair No (4) Replaced the New dust cover, three way syringe and 2 nos. air tube (ig) against old. Flow checkers and found in the working condition.

attended the complaint in the chair No. air tube has damaged. Replaced the new air tube in 2 nos (ig) against old.

Spares - No tube
Three way syringe - 10.
Air tube holder
Dust cover


PRINCIPAL
The Oxford Dental College,
Bommanahalli, Hosur Road,
Bangalore - 560 068.

